

Certificate

Traceability of Manufacturers to the IFCC Reference Measurement Procedure for HbA1c

This certifies that Bayer Diabetes Care using AICNow+, uses calibrators supplied by the IFCC Network to get traceable to the IFCC Reference Measurement Procedure and participates in the Monitoring Programme to demonstrate traceability. In the Monitoring Programme of 2013 the following performance was seen:

Deviation from IFCC-target	at 30 mmol HbA1c/mol Hb :	-3.9
	at 60 mmol HbA1c/mol Hb :	-4.1
	at 90 mmol HbA1c/mol Hb :	-4.3
Reproducibility, coefficient of variation		7.16%
Linearity, correlation coefficient		0.9828

* Bayer HealthCare Addendum: See attached statement regarding commutability of IFCC traceability samples

Date of issue: 5 December 2013

Certification expires: 31 December 2014



IFCC Network Coordinator