

PTS Panels® HDL Cholesterol Control Ranges		Level 1		Level 2	
Lot #:	HC23	mg/dL	mmol/L	mg/dL	mmol/L
Expiration Date:	2017 07 31				
Single Analyte Test Strips					
HDL Cholesterol		≤25 - 45	≤0.65 - 1.17	30 - 60	0.78 - 1.55
Multi-Analyte Test Strips					
HDL Cholesterol		15 - 50	0.39 - 1.30	30 - 70	0.78 - 1.81
<p>For assistance, contact PTS Customer Service @ 877-870-5610 or Fax @ 317-870-5608. Outside of US @ 317-870-5610. Email: inforequest@cardiochek.com</p>					
PS-002516 Rev. 6					

PTS Panels® HDL Cholesterol Control Ranges		Level 1		Level 2	
Lot #:	HC23	mg/dL	mmol/L	mg/dL	mmol/L
Expiration Date:	2017 07 31				
Single Analyte Test Strips					
HDL Cholesterol		≤25 - 45	≤0.65 - 1.17	30 - 60	0.78 - 1.55
Multi-Analyte Test Strips					
HDL Cholesterol		15 - 50	0.39 - 1.30	30 - 70	0.78 - 1.81
<p>For assistance, contact PTS Customer Service @ 877-870-5610 or Fax @ 317-870-5608. Outside of US @ 317-870-5610. Email: inforequest@cardiochek.com</p>					
PS-002516 Rev. 6					

PTS Panels® HDL Cholesterol Control Ranges		Level 1		Level 2	
Lot #:	HC23	mg/dL	mmol/L	mg/dL	mmol/L
Expiration Date:	2017 07 31				
Single Analyte Test Strips					
HDL Cholesterol		≤25 - 45	≤0.65 - 1.17	30 - 60	0.78 - 1.55
Multi-Analyte Test Strips					
HDL Cholesterol		15 - 50	0.39 - 1.30	30 - 70	0.78 - 1.81
<p>For assistance, contact PTS Customer Service @ 877-870-5610 or Fax @ 317-870-5608. Outside of US @ 317-870-5610. Email: inforequest@cardiochek.com</p>					
PS-002516 Rev. 6					